



2525 Tollgate Rd, Quakertown, PA 18951
 Phone: 215-536-4400 Fax: 215-536-4096
 www.precisionsolutionsinc.com

Electronic Data Report

Report #: 38643-001268-04
 Service Date: 05-Aug-2021
 Technician: Trent Bradford

Customer Information

Client: Customer Name Inc
 123 Street Road
 Quakertown, PA 18951

Location of Calibration: Customer Name Inc
 123 Street Road
 Quakertown, PA 18951

Asset Information

Manufacturer: Ohaus
Model: ARC120
Serial #: J04351213
Customer ID: QC
Location: QC
Range:
Capacity: 3100 g
Resolution: 0.01 g

As Found Readings

Applied	Reading
Zero (g)	
0.00	0.00
Increasing Load (g)	
100.00	99.98
300.00	299.93
1000.00	999.85
2000.00	1999.74
3000.00	2999.77
Shift Round (g)	
500.00	499.89
500.00	499.87
500.00	499.91

As Left Readings

Applied	Reading
Zero (g)	
0.00	0.00
Increasing Load (g)	
100.00	100.00
300.00	300.00
1000.00	1000.00
2000.00	1999.95
3000.00	3000.00
Shift Round (g)	
500.00	499.99
500.00	499.99
500.00	500.00

Reference Standard(s) Used ID:

Set #4

Comments:

Inspected, tested, and calibrated